

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

JUL 24 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130172</u> ✓	
Name <u>Brent Johnson</u>		ID# <u>4980</u>	Date <u>7/18/09</u> ✓
A	Agency <u>AWT</u>	Phone # <u>883-5111</u>	
Instrument Location <u>Tok</u>		✓	
Alco S/N <u>X173017</u> ✓		Target Value <u>.083</u> ✓	High Pressure <u>900</u>
B	Alco Test Value Average <u>.083</u> ✓		
	1 st Alco	2 nd Alco	
Signature <u>[Signature]</u>		<u>COB</u> <u>7/24/09</u>	
(OVER)			

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 7th day of Aug, 2009.

[Signature]
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130172

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130172

JULY 18, 2009

OPERATOR'S NAME:

JOHNSON BRENT W

OPERATOR'S NUMBER: 4900

SUBJECT'S LAST NAME:

VERIF

SUBJECT'S FIRST NAME/MI :

OF CALIB

O.L. #: 000000

DEPT/AGENCY: TOKO

CASE/REPORT: 00-000

TEST TYPE: U

ALCO TARGET VALUE: .083 ✓

ALCO S/N: X173017 ✓

— BREATH ANALYSIS —

ALCO ADJUSTED FOR 29.34 in		13:00
ALCO TARGET	.083	12:59
BLANK TEST	.000	12:59
INTERNAL STANDARD	VERIFIED	12:59
ALCO TO 29.34 in	.083 ✓	12:59
BLANK TEST	.000	13:00
SUBJECT SAMPLE	.083 ✓	13:01
BLANK TEST	.000	13:01
ALCO TO 29.34 in	.083 ✓	13:02
BLANK TEST	.000	13:03

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130172

JULY 18, 2009

TIME 13:03

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.34 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

"002"(C)++,-./0123456789::(<=>?@ABCDEF6
HIJKLMNOPQRSTUVWXYZ\|_! "abcde fghijklmno
pqrstuvwxyz{|}~■